



Certificate of Analysis

QC Code: GIBCO

Distilled Water

Lot Number: 3142878
Item Number: 15230
Expiration Date: 2026-03
Storage Temp: 15 to 30C

For research use only. Not for use in diagnostic procedures.
Sterile filtered (0.1 um)

TEST	TEST ID	SPECIFICATION	RESULT	UNITS
Endotoxin Testing	ENDO0007	Check and Record	<0.01	EU/mL
Osmolality	OSMO0002	0 - 5	4	mOsm/kg
pH	PH0003	5.0 - 7.5	5.8	
Sterility Testing	STERI0007	Negative	Negative	

Read SDS

Quality Systems Department

Date: 10-Apr-2025

References

- ENDO0007: Current United States Pharmacopeia, <85> Bacterial Endotoxins Test.
- OSMO0002: Thermo Fisher Scientific Specifications.
- PH0003: Thermo Fisher Scientific Specifications.
- STERI0007: Current Edition of USP, Thermo Fisher Scientific Modified.